

Medaval Accreditation Assessment

Volume 2016

Report 1601

05 August 2016

Accreditation assessment of the blood pressure measurement technology used in the iHealth BP3 upper arm monitor, as validated according to the European Society of Hypertension International Protocol revision 2010

Approved by the Medaval Advisory Board

Eoin O'Brien (Chair)	Yutaka Imai	Andrew Shennan
George S. Stergiou (Deputy Chair)	Martin Myers	Jan Staessen
Roland Asmar	Gbenga Ogedegbe	Martin J. Turner
Alejandro de la Sierra	Takayoshi Ohkubo	Paolo Verdecchia
Peter W. de Leeuw	Paolo Palatini	Bernard Waeber
Eamon Dolan	Gianfranco Parati	J-Guang Wang
Geoffrey A. Head		

Reference

Medaval Ltd. Accreditation assessment of the blood pressure measurement technology used in the iHealth BP3 upper arm monitor, as validated according to the European Society of Hypertension International Protocol revision 2010. *Medical Device Assessment*. 2016 Aug 5; **2016**(1601). 5 p. Epub: 2019 Jan 30. Available from: <https://www.medaval.ie/MDA/2016/MDA1601.pdf>.

Medical Device Assessment is published by

Medaval Ltd., Unit 107, SBC, Serpentine Ave., Ballsbridge, Dublin D04 H522, IRELAND.


© 2016-2019 Medaval Ltd. All rights reserved.

Permissions: Requests for permissions to reproduce figures, tables, or portions of reports or articles originally published in *Medical Device Assessment* can be obtained by email request to info@medaval.ie.

Accreditation assessment of the blood pressure measurement technology used in the iHealth BP3 upper arm monitor, as validated according to the European Society of Hypertension International Protocol revision 2010

Medaval Accreditation-Assessment Report – 5th August 2016

Test Device Details

		Assessment
Full Name	iHealth BP3	Requirement satisfactory
Model	BP3	Requirement satisfactory
Measurement Site	Upper Arm	Requirement satisfactory
Client Use	Suitable for self-measurement.	Requirement satisfactory
Operation Method	Oscillometry, automatic during deflation	Requirement satisfactory
Measurement Occurrence	Single Measurements Only	Requirement satisfactory
Device Photograph		Requirement satisfactory
Manufacturer(s)	OEM: Andon Health Co. Ltd., 3 Jinping Road, Ya'an Street, Nankai District, Tianjin 300190, CHINA OBL: iHealth Lab Inc, 3 rue Tronchet, 75008 Paris, FRANCE	Requirement satisfactory
Cuffs	Standard: 22 cm to-30 cm Large: 30 cm to-42 cm XLarge: 42 cm to-48 cm	Cuffs Listed: Requirement satisfactory Arm Circumferences: Requirement satisfactory

Study Details

Original Publication	Chen C, Shang F, Wang J, Chen J, Ji N, Wan Y. Validation of the iHealth BP3 upper-arm blood pressure monitor, for clinic use and self-measurement, according to the European Society of Hypertension International Protocol revision 2010. <i>Blood Press Monit.</i> 2012 Dec; 17 (6):253-6. doi: 10.1097/MBP.0b013e32835b9ecb. PMID: 23147535.
Protocol	The European Society of Hypertension International Protocol revision 2010 for the validation of blood pressure measuring devices in adults ¹

		Assessment
Adherence	Followed Precisely	Requirement satisfactory
Adjustments	None	Requirement satisfactory
Study Meas. Method	Oscillometric	Requirement satisfactory
Study Measurement Site	Upper Arm	Requirement satisfactory
Observers		
Supervisor + 2 Observers	Yes	Requirement satisfactory
Observer Training	Not stated.	Modification: Inferred from following protocol precisely.
Observer Familiarisation	12 test measurements	Requirement satisfactory
Observers Blinding	From device and each other	Requirement satisfactory
Sample		
Population	A general population	Requirement satisfactory
Circumstances	None	Requirement satisfactory
HBP Subjects Selection	Inpatients and outpatients	Requirement satisfactory
NBP Subjects Selection	Accompanying persons & staff	Requirement satisfactory

Test Device Details and Study Details Assessment	Checks	22
	Permitted Modifications	1
	Violations	0

Procedure

Table 1: Screening and Recruitment Details

Screening and Recruitment				Assessment
Total Screened			54	Value within requirements
Total Excluded			21	Value within requirements
	Ranges Complete		16	Value within requirements
	Range Adjustment		5	Value within requirements
	Arrhythmias		0	Value within requirements
	Device Failure		0	Value within requirements
	Poor Quality Sounds		0	Value within requirements
	Cuff Size Unavailable		0	Value within requirements
	Observer Disagreement		0	Value within requirements
	Distribution		0	Value within requirements
	Other Reasons*		0	Value within requirements
Total Recruited			33	Value within requirements
*Explanation Summary				No details required
Recruitment Ranges				
SBP	Total		33	Value within requirements
	Low	< 90 mmHg	0	Value within requirements
		90 – 129 mmHg	11	Value within requirements
		130 – 160 mmHg	11	Value within requirements
	Medium	161 – 180 mmHg	10	Value within requirements
		> 180 mmHg	1	Value within requirements
	DBP	Total		33
Low		< 40 mmHg	0	Value within requirements
		40 – 79 mmHg	11	Value within requirements
		80 – 100 mmHg	10	Value within requirements
High		101 – 130 mmHg	12	Value within requirements
		> 130 mmHg	0	Value within requirements
Total Extremes			1	Value within requirements
On Treatment Ranges				
SBP	Low	< 130 mmHg	1	Value within requirements
	Medium	130 – 160 mmHg	5	Value within requirements
	High	> 160 mmHg	7	Value within requirements
DBP	Low	< 80 mmHg	3	Value within requirements
	Medium	80 – 100 mmHg	3	Value within requirements
	High	> 100 mmHg	7	Value within requirements

Table 1 Assessment	Checks	36
	Permitted Modifications	0
	Violations	0

Study Results

Table 2: Subject Details

			<u>Assessment</u>	
Sex	Male:Female	18:15	Value within requirements	Value within requirements
Age (years)	Range (Low:High)	27:69	Value within requirements	Value within requirements
	Mean (SD)	47.1 (12.3)	Value within requirements	Value within requirements
Arm Circumference (cm)	Range (Low:High)	22:38	Value within requirements	Value within requirements
	Mean (SD)	30.0 (4.4)	Value within requirements	Value within requirements
Cuff for Test Device (cm)	Standard (22 to 30)	17		
	Large (30 to 42)	16		
	XLarge (42 to 48)	0		
	Total	33	Value within requirements	
Recruitment SBP (mmHg)	Range (Low:High)	89:190	Value within requirements	Value within requirements
	Mean (SD)	143.9 (27.4)	Value within requirements	Value within requirements
Recruitment DBP (mmHg)	Range (Low:High)	46:121	Value within requirements	Value within requirements
	Mean (SD)	90.1 (18.3)	Value within requirements	Value within requirements
Table 2 Assessment			Checks	19
			Permitted Modifications	0
			Violations	0

Table 3: Observer Measurements in each Recruitment Range

			<u>Assessment</u>	
SBP	Overall Range mmHg (Low:High)	91:192	Value within requirements	Value within requirements
	Low (< 130 mmHg)	31		Value within requirements
	Medium (130 – 160 mmHg)	35		Value within requirements
	High (> 160 mmHg)	33		Value within requirements
	Maximum Difference	4		Value within requirements
DBP	Overall Range mmHg (Low:High)	46:122	Value within requirements	Value within requirements
	Low (< 80 mmHg)	29		Value within requirements
	Medium (80 – 100 mmHg)	34		Value within requirements
	High (> 100 mmHg)	36		Value within requirements
	Maximum Difference	7		Value within requirements
Table 3 Assessment			Checks	12
			Permitted Modifications	0
			Violations	0

Table 4: Observer Differences



			<u>Assessment</u>	
Observer 2 – Observer 1				
SBP (mmHg)	Range (Low:High)	-4:+4	Value within requirements	Value within requirements
	Mean (SD)	+0.1 (2.3)	Value within requirements	Value within requirements
DBP (mmHg)	Range (Low:High)	-4:+4	Value within requirements	Value within requirements
	Mean (SD)	+0.2 (2.0)	Value within requirements	Value within requirements
Repeated Measurements		3	Value within requirements	
Table 4 Assessment			Checks	9
			Permitted Modifications	0
			Violations	0

Table 5: Validation Results

Part 1	Pass Req.		Achieved		Assessment	
	Two of	All of	SBP	DBP		
≤ 5 mmHg	73	65	72	85	Value within lower passing criteria	Value within passing criteria
≤ 10 mmHg	87	81	94	99	Value within passing criteria	Value within passing criteria
≤ 15 mmHg	96	93	99	99	Value within passing criteria	Value within passing criteria
Grade 1			Pass	Pass	Value within lower passing criteria	Value within passing criteria
Mean mmHg			2.8	-0.4	Value within requirements	Value within requirements
SD mmHg			4.2	3.5	Value within requirements	Value within requirements
Part 2	Pass Req.	Achieved	SBP	DBP		
2/3 ≤ 5 mmHg	≥ 24	24	29		Value within passing criteria	Value within passing criteria
0/3 ≤ 5 mmHg	≤ 3	0	1		Value within passing criteria	Value within passing criteria
Grade 2		Pass	Pass		Value within passing criteria	Value within passing criteria
Grade 3		Pass	Pass		Value within lower passing criteria	Value within passing criteria
Part 3	Result	Pass	Value within lower passing criteria			

Table 5 Assessment	Checks	21
	Permitted Modifications	0
	Violations	0

Plots



SBP Plot Provided **Yes**
 DBP Plot Provided **Yes**

Assessment	
Requirement satisfactory	
Requirement satisfactory	

Plots Assessment	Checks	2
	Permitted Modifications	0
	Violations	0

Recommendations

Overall Summary

Number of checks 121
 Number of permitted modifications 1
 Number of violations 0

Assessment Summary

The validation has been checked and is verified as having been conducted in accordance with the protocol requirements. Therefore, the results are considered to be valid, the null hypothesis, that the device is inaccurate in measuring blood pressure, is rejected and the conclusion, that the device is accurate for self-measurement in adults, is correct.

Certification Decision

The iHealth BP3, with either the Standard 22 cm to-30 cm or the Large 30 cm to-42 cm cuff, is certified by Medaval Ltd., for blood pressure measurement in adults, as it fulfilled the conditions required for a pass in a validation study carried out in accordance with the requirements of the International Protocol of the European Society of Hypertension 2010 Revision.

Date of Advisory Board Approval: 3rd August 2016.

Reference

- O'Brien E, Atkins N, Stergiou G, Karpettas N, Parati G, Asmar R, Imai Y, Wang J, Mengden T, Shennan A; Working Group on Blood Pressure Monitoring of the European Society of Hypertension. European Society of Hypertension International Protocol revision 2010 for the validation of blood pressure measuring devices in adults. *Blood Press Monit.* 2010;**15**:23-38. doi: 10.1097/MBP.0b013e3283360e98. PMID: 20110786. Erratum in *Blood Press Monit.* 2010;**15**(3):171-2.