

## Medaval Accreditation Assessment

Volume 2019

Report 1902AR

05 March 2019

**Accreditation assessment of the auscultatory blood pressure measurement technology used in the Novacor Diasys 3 Plus (DIP-0001-00) upper arm ABPM monitor, as validated according to the AAMI/ANSI/ISO 81060-2:2013 standard both for a general study in adults and for an additional ABPM-device study and according to the European Society of Hypertension International Protocol revision 2010**

<b>Investigators</b>	Paolo Palatini, Studioso senior dello Studium Patavinum, University of Padova, ITALY. Claudio Fania, Reparto di Medicina, Casa di Cura Villa Maria SRL., Padova, ITALY.
<b>Analysis</b>	Neil Atkins, Medaval Ltd., Dublin, IRELAND.
<b>Reference</b>	Medaval Ltd. Accreditation assessment of the auscultatory blood pressure measurement technology used in the Novacor Diasys 3 Plus (DIP-0001-00) upper arm ABPM monitor, as validated according to the AAMI/ANSI/ISO 81060-2:2013 standard both for a general study in adults and for an additional ABPM-device study and according to the European Society of Hypertension International Protocol revision 2010. <i>Medical Device Assessment</i> . 2019 Mar 05;2019(1902AR) 23 p. Available from: <a href="https://www.medaval.ie/MDA/2019/MDA1902AR.pdf">https://www.medaval.ie/MDA/2019/MDA1902AR.pdf</a> .

*Medical Device Assessment* is published by

Medaval Ltd., Unit 107, SBC, Serpentine Ave., Ballsbridge, Dublin D04 H522, IRELAND.

© 2019 Medaval Ltd. All rights reserved.

**Permissions:** Requests for permissions to reproduce figures, tables, or portions of reports or articles originally published in *Medical Device Assessment* can be obtained by email request to [info@medaval.ie](mailto:info@medaval.ie).

## MEDICAL DEVICE ASSESSMENT 1902AR:2019

# Accreditation assessment of the auscultatory blood pressure measurement technology used in the Novacor Diasys 3 Plus (DIP-0001-00) upper arm ABPM monitor, as validated according to the AAMI/ANSI/ISO 81060-2:2013 standard both for a general study in adults and for an additional ABPM-device study and according to the European Society of Hypertension International Protocol revision 2010

*Medaval Accreditation Assessment – 05 March 2019*

## Summary

### Introduction

The Novacor Diasys 3 Plus (model number DIP-0001-00) is a blood pressure monitor intended for ambulatory blood pressure measurement. Measurements are recorded in hybrid mode with auscultation as the primary method and a backup using oscillometry. They are recorded automatically at timed intervals or additionally on indication of postural hypotension.

### Methodology

Under quiet laboratory conditions, when the Novacor Diasys 3 Plus was set to hybrid mode, it was able to detect all measurements using auscultation and, in this way, the oscillometric option was excluded effectively.

The Novacor Diasys 3 Plus was validated, using auscultatory measurements only, according to the requirements of the AAMI/ANSI/ISO 81060-2:2013 standard<sup>1</sup>. This standard requires two validations for ABPM devices: The first is a primary study in a sample of at least 85 subjects from a general population, in accordance with specified recruitment requirements. The second, an additional requirements for ABPM devices, is a study in a sample of at least 35 subjects who's pulse rate is increased by between 10% and 20% by exercising on bicycle ergometer in order to achieve this change. The Novacor Diasys 3 Plus was also validated according to the requirements of the European Society of Hypertension International Protocol revision 2010 (ESH-IP)<sup>2</sup>. This is a study in a sample of 33 subjects from a general population, in accordance with very strict recruitment requirements.

A total of 85 subjects were included. Thirty-three of the first 35 subjects comprised the sample required for the ESH-IP study. Thirty six of the remaining 50 subjects also comprised the sample required for the additional requirements study. No subject who participated in the ESH-IP study also participated in the additional requirements study.

Three cuffs were used in accordance with the requirements of the protocols. These were the Paediatric Plus 18 cm to 24 cm cuff (model number ACC-0212-00), the Standard Plus 24 cm to 32 cm cuff (model number ACC-0210-00) and the Large Plus 32 cm to 40 cm cuff (model number ACC-0211-00).

### Results

All of the requirements of each of the protocols were satisfied without any adjustments or violations.

In the primary AAMI/ANSI/ISO 81060-2:2013 study, the Criterion 1 errors were  $+3.6 \text{ mmHg} \pm 2.7 \text{ mmHg}$  for SBP and  $+3.0 \pm 2.7 \text{ mmHg}$  for DBP and all values were within the  $\pm 5.0 \text{ mmHg} \pm 8.0 \text{ mmHg}$  requirements. The Criterion 2 errors were  $+3.6 \text{ mmHg} \pm 2.6 \text{ mmHg}$  for SBP (within the  $\pm 5.0 \text{ mmHg} \pm 5.89 \text{ mmHg}$  requirements) and  $+3.0 \pm 2.5 \text{ mmHg}$  for DBP (within the  $\pm 5.0 \text{ mmHg} \pm 6.25 \text{ mmHg}$  requirements).

In the additional requirements AAMI/ANSI/ISO 81060-2:2013 study, the Criterion 1 errors were  $+4.0 \text{ mmHg} \pm 3.3 \text{ mmHg}$  for SBP and  $+4.0 \pm 3.4 \text{ mmHg}$  for DBP and all values were within the  $\pm 5.0 \text{ mmHg} \pm 8.0 \text{ mmHg}$  requirements. The Criterion 2 errors were  $+4.0 \text{ mmHg} \pm 3.1 \text{ mmHg}$  for SBP (within the  $\pm 5.0 \text{ mmHg} \pm 5.64 \text{ mmHg}$  requirements) and  $+4.0 \pm 3.3 \text{ mmHg}$  for DBP

(within the  $\pm 5.0$  mmHg  $\pm 5.64$  mmHg requirements).

The ESH-IP protocol has two set of passing requirements. Part 1 sets out ideal and minimal accuracy levels for the 99 individual pairs of measurements each with three criteria. At least two of the ideal criteria and all three of the minimal criteria must be met. The ideal criteria are at least 73 differences (between test and reference measurement) are within 5 mmHg, at least 87 differences are within 10 mmHg and at least 96 differences are within 15 mmHg. For the Novacor Diasys 3 Plus, those differences were 94, 99 and 99 respectively for SBP and were also 94, 99 and 99 respectively for DBP. Part 1 criteria were, therefore, met for both SBP and DBP. Part 2 sets out accuracy levels for the sets of three measurements from the 33 subjects. At least 24 subjects must have at least two measurements within 5 mmHg of the reference measurements and at most 3 subjects can have no measurements within 5 mmHg of the reference measurements. For the Novacor Diasys 3, those counts were 32 and 0 subjects respectively for SBP and were 33 and 0 subjects respectively for DBP. Part 2 criteria were, therefore, met for both SBP and DBP.

## Conclusion

Each of the protocols are designed to test the null hypothesis that the device is inaccurate and that this hypothesis must be rejected if the respective passing criteria are met.

As the protocols were followed strictly, any hypothesis that the reliability of the results may be compromised due to protocol adjustment or violation must also be rejected.

Therefore, as the passing criteria in each case have been met, there is no option but to reject the null hypothesis and conclude that the auscultatory blood pressure measurement technology used in the Novacor Diasys 3 Plus is accurate, when used with the appropriate cuff, as described herein, within the criteria set out in AAMI/ANSI/ISO 81060-2:2013 for both a primary study in adults and an additional requirements for ABPM study and also within the criteria set out in the European Society of Hypertension International Protocol revision 2010 for a study in a general population.

Due to dual recruitment, neither the ESH-IP or AAMI/ANSI/ISO 81060-2:2013 additional requirements study can be considered as being entirely independent of the AAMI/ANSI/ISO 81060-2:2013 primary study. However, the ESH-IP and AAMI/ANSI/ISO 81060-2:2013 additional requirements studies are mutually independent.

## Contents

<b>SUMMARY .....</b>	<b>2</b>
<i>Introduction.....</i>	<i>2</i>
<i>Methodology.....</i>	<i>2</i>
<i>Results.....</i>	<i>2</i>
<i>Conclusion .....</i>	<i>3</i>
<b>GLOSSARY.....</b>	<b>6</b>
<i>Abbreviations.....</i>	<i>6</i>
<i>Plot Legend.....</i>	<i>6</i>
<b>ORGANISATIONAL DETAILS .....</b>	<b>7</b>
<i>Medaval Ltd .....</i>	<i>7</i>
<i>Casa di Cura Villa Maria SRL .....</i>	<i>7</i>
<i>Novacor.....</i>	<i>8</i>
<b>DEVICE DETAILS .....</b>	<b>9</b>
TEST DEVICE.....	9
REFERENCE SPHYGMOMANOMETER.....	9
REFERENCE CUFF(S).....	9
REFERENCE STETHOSCOPE .....	9
<b>AAMI/ANSI/ISO 81060-2:2013 PRIMARY STUDY .....</b>	<b>10</b>
STUDY DETAILS.....	10
<i>Observers.....</i>	<i>10</i>
<i>Sample.....</i>	<i>10</i>
<i>Subject Preparation .....</i>	<i>10</i>
PROCEDURE.....	10
<i>Screening and Recruitment Details .....</i>	<i>10</i>
<i>Subject Details.....</i>	<i>11</i>
<i>Observer Measurements Range-Requirements .....</i>	<i>11</i>
STUDY RESULTS .....	12
<i>Observer Differences.....</i>	<i>12</i>
<i>Validation Results .....</i>	<i>12</i>
<i>Plots.....</i>	<i>12</i>
<b>AAMI/ANSI/ISO 81060-2:2013 ADDITIONAL REQUIREMENTS FOR ABPM DEVICE .....</b>	<b>14</b>
STUDY DETAILS.....	14
<i>Observers.....</i>	<i>14</i>
<i>Sample.....</i>	<i>14</i>
PROCEDURE.....	14
<i>Screening and Recruitment Details .....</i>	<i>14</i>
<i>Subject Details.....</i>	<i>15</i>
<i>Stress Details .....</i>	<i>15</i>
<i>Observer Measurements Range-Requirements .....</i>	<i>16</i>
STUDY RESULTS .....	16
<i>Observer Differences.....</i>	<i>16</i>
<i>Validation Results .....</i>	<i>16</i>
<i>Plots.....</i>	<i>17</i>
<b>ESH-IP 2010 STUDY .....</b>	<b>18</b>
STUDY DETAILS.....	18
<i>Observers.....</i>	<i>18</i>
<i>Sample.....</i>	<i>18</i>
<i>Subject Preparation .....</i>	<i>18</i>

PROCEDURE..... 19  
*Table 1: Screening and Recruitment Details*..... 19  
*Table 2: Subject Details* ..... 20  
*Table 3: Observer Measurements in each Recruitment Range*..... 20  
 STUDY RESULTS ..... 21  
*Table 4: Observer Differences* ..... 21  
*Table 5: Validation Results*..... 21  
*Plots*..... 22  
**LIMITATIONS .....23**  
**RECOMMENDATIONS.....23**  
**REFERENCES .....23**

## Glossary

### Abbreviations

%RH	relative humidity percent
°C	degrees Celsius
AAMI	Association for the Advancement of Medical Instrumentation (USA)
ABPM	Ambulatory blood pressure measurement
AMP	Associazione Medica Patavina (Medical Association Patavina – Italy)
ANSI	American National Standards Institute
BP	Blood pressure
bpm	beats per minute
CE	Conformité Européene (European Conformity)
CEE	Communauté Economique Européenne (European Economic Community – former name for EU)
cm	centimetre(s)
DBP	Diastolic blood pressure
DIMED	Dipartimento di Medicina (Department of Medicine)
ECG	Electrocardiogram
ESH	European Society of Hypertension
ESH-IP	European Society of Hypertension International Protocol
EU	European Union
g	gram(s)
hPa	hectopascals
ISO	International Organization for Standardization/International Standards Organization
m	metre(s)
mAh	milliamp hours
MAP	Mean arterial pressure
meas.	Measurement
min	minute(s)
mmHg	millimetre(s) of mercury
NiMH	Nickel–metal hydride (battery)
PC	Personal computer (any external system do which data can be downloaded)
PP	Pulse Pressure (SBP – DBP)
PR	Pulse rate
QKD & QKd	The QKD interval is the time between an ECG QRS wave and the last Korotkoff sound during BP measurement.
req.	requirement
SBP	Systolic blood pressure
SD	standard deviation
SIIA	Società Italiana per l'Ipertensione Arteriosa (Italian Society for Arterial Hypertension)
SRL	Società a responsabilità limitata (Italian equivalent to a limited liability company)
VR	Verona (Italy)

### Plot Legend

- Single point, area one unit
- Two superimposed points, area two units
- Three superimposed points, area two units
- Four superimposed points, area four units
- Five superimposed points, area five units

## Organisational Details

### Medaval Ltd.

Incorporated in 1989 as Medical Device Assessment Ltd, the company abbreviated its name to Medaval Ltd. in 2015. Medaval provides several services including comprehensive cardiovascular device listings according to peer-reviewed validations, certification for devices that have been proven to have been validated strictly according to a current standard protocol, validation of devices and comparative-equivalence according to MEDDEV 2.7/1 rev 4 standards<sup>3</sup>. Both validation and comparative-equivalence services are in accordance with Regulation (EU) 2017/745.

Validations are performed by members of a Validation Panel. They are assigned blind to the manufacturer. Validation reports are peer reviewed by members from a corresponding Reviewer Panel. Once reviewed, investigators are free to prepare a scientific paper for publication, should they so wish.

The passing criteria in validation protocols are based on specific sample distributions and on other criteria and can only be applied if all of the requirements are followed correctly. Therefore, in any validation study, Medaval, first tests the hypothesis that the study was not carried out in accordance with the requirements and it is only if that hypothesis is rejected can the results be considered reliable.

All procedures are developed and reviewed by members of our Scientific Procedures Panel. The Medaval Accreditation Procedure is designed to check that every aspect of a validation protocol is fulfilled. Modifications, that may be necessary for particular populations or circumstances not defined specifically in a protocol must be supported by relevant peer-reviewed scientific publications.

Validation is considered to apply to the specific measurement technology being tested, as distinct from the device itself. No inference should be made about the validity of any other aspect of the device, unless it is also tested according to a regulatory or peer-reviewed protocol. Validation also only applies to the population from which the sample is taken and under the circumstances in which it was carried out, as defined in the protocol. No inference should be made about the validity of the device in a different population or under different circumstances.

The results must apply equally to any device that uses an equivalent measurement technology, as proven under MEDDEV 2.7/1 rev 4 standards irrespective of whether that equivalence is proven prior to or subsequent to the validation. The Medaval Scientific Procedures Panel has developed as comparative-equivalence procedure to test the null hypothesis that two devices are not equivalent, according to this standard. Should that hypothesis be rejected, the devices must be regarded as equivalent for that measurement technology.

For more information, please refer to [www.medaval.ie](http://www.medaval.ie).

### Casa di Cura Villa Maria SRL

The study investigators were

Prof. Paolo Palatini, Department of Medicine, University of Padova, ITALY and .

Dr. Claudio Fania, Reparto di Medicina, Casa di Cura Villa Maria SRL., Padova, ITALY.

Casa di Cura Villa Maria is one of the most important hospitals, accredited within the Italian National Health System, in Padua.

It has departments in Orthopaedic rehabilitation, Neurological rehabilitation, Rehabilitative long-term care, Plastic surgery and General medicine as well as extensive and intensive outpatient services.

Dr. Claudio Fania is a researcher and an expert in hypertension, on early diagnosis, in the study of possible secondary causes, in the study and prevention of organ damage, in appropriate therapy and in the validation of blood pressure monitors. He worked for years, alongside Prof. Paolo Palatini, at the SIIA-certified hypertension centre at the DIMED Department (Department of Medicine) of the University of Padua and now works in the General Medicine Department at Casa di Cura Villa Maria.

He is the author of 40 international publications in the fields of hypertension, cardiovascular diseases and blood pressure monitor validation. He is a member of the SIIA (Italian Society for Arterial Hypertension), the ESH (European Society of Hypertension) and the AMP (Medical Association Patavina).

For more information, please refer to [www.cdcvillamaria.it/pag/casa-di-cura-villa-maria/](http://www.cdcvillamaria.it/pag/casa-di-cura-villa-maria/)

**Novacor**

Novacor is one of the world leaders in ambulatory recording. It has been operating in this field successfully for more than 25 years and offers a complete and advanced range of ECG Recorders and Ambulatory Blood Pressure Monitors.

Novacor co-operates with healthcare professionals to improve diagnostic devices continuously to meet clinical needs more closely whilst ensuring

each product remains simple to use for the specialist and convenient for the patient. Novacor provides patients and healthcare professionals with products that combine Innovation, Reliability and Quality.


The company has the ISO 13485 certificates. All the products have the CE mark (Dir 93/42 CEE).

For more information, please refer to [www.novacor.com](http://www.novacor.com).



## Device Details

### TEST DEVICE

		<u>Assessment</u>
<b>Full Name</b>	Novacor Diasys 3 Plus	Requirement satisfactory
<b>Model</b>	DIP-0001-00	Requirement satisfactory
<b>Measurement Site</b>	Upper Arm	Requirement satisfactory
<b>Client Use</b>	Suitable for ABPM.	Requirement satisfactory
<b>Operation Method</b>	1: Hybrid (Auscultatory with Oscillometric backup), automatic during deflation 2: Oscillometry, automatic during deflation	Requirement satisfactory
<b>Measurement Occurrence</b>	Single Measurements Only	Requirement satisfactory
<b>Device Photograph</b>		Requirement satisfactory
<b>Manufacturer(s)</b>	NOVACOR 4 Passage Saint-Antoine, 92508 Rueil-Malmaison, FRANCE	Requirement satisfactory
<b>Cuffs</b>	Paediatric Plus (ACC-0212-00): 18 cm to 24 cm Standard Plus (ACC-0210-00): 24 cm to 32 cm Large Plus (ACC-0211-00): 32 cm to 40 cm	Cuffs Listed: Requirement satisfactory Arm Circumferences: Requirement satisfactory

### REFERENCE SPHYGMOMANOMETER

		<u>Assessment</u>
<b>Full Name</b>	BoSo Mercurius E	Information requirement
<b>Model</b>	370-0-101	Information requirement
<b>ISO-81060-1 Certification</b>	Yes	Information requirement
<b>Manufacturer(s)</b>	Bosch + Sohn GmbH u. Co. KG, Bahnhofstraße 64, D-72417 Jungingen, GERMANY.	Desirable information

### REFERENCE CUFF(S)

		<u>Assessment</u>
<b>Name(s) and Model(s)</b>	Bracelet for the measurement of PA for sphygmomanometers (SBEC-2U) 14 cm to 22 cm (SBEC-2U) 22 cm to 32 cm (SBEC-2U) 32 cm to 42 cm	Cuffs Listed: Information requirement Arm Circumferences: Information requirement
<b>ISO-81060-1 Certification</b>	Yes	Information requirement
<b>Manufacturer(s)</b>	MediCare - El.Med. Garda srl Via San Giuseppe Artigiano 6, 37010 Costermano sul Garda VR, ITALY	Desirable information

### REFERENCE STETHOSCOPE

		<u>Assessment</u>
<b>Full Name</b>	Classic Teaching Stetho (for dual simultaneous auscultation)	Desirable information
<b>Model</b>	32542	Desirable information
<b>Manufacturer(s)</b>	Gima S.p.A., Via Marconi 1, 20060 Gessate Milano, ITALY.	Desirable information

<b>Device Details Assessment</b>	<b>Checks</b>	21
	<b>Permitted Modifications</b>	0
	<b>Violations</b>	0

# AAMI/ANSI/ISO 81060-2:2013 Primary Study

## STUDY DETAILS

<b>Protocol</b>	The AAMI/ANSI/ISO 81060-2:2013 standard for a general study in adults <sup>1</sup>	
		<b>Assessment</b>
<b>Reference Determination</b>	Sequential same-arm	Information requirement satisfactory
<b>Adherence</b>	Followed Precisely	Information requirement satisfactory
<b>Adjustments</b>	None	Information requirement satisfactory
<b>Study Meas. Method</b>	Auscultatory	Information requirement satisfactory
<b>Study Measurement Site</b>	Upper Arm	Information requirement satisfactory
<b>Observers</b>		
<b>Supervisor + 2 Observers</b>	Yes	Information requirement satisfactory
<b>Observer Training</b>	By expert in BP measurement	Information requirement satisfactory
<b>Observer Familiarisation</b>	40 test measurements	Supplementary Information
<b>Observers Blinded</b>	From device and each other	Information requirement satisfactory
<b>Sample</b>		
<b>Population</b>	A general population	Information requirement satisfactory
<b>Circumstances</b>	None	Information requirement satisfactory
<b>HBP Subjects Selection</b>	Inpatients and outpatients	Supplementary Information
<b>NBP Subjects Selection</b>	Healthcare staff	Supplementary Information
<b>Subject Preparation</b>		
<b>Back, elbow and arm supported</b>	Yes	Information requirement
<b>Legs uncrossed</b>	Yes	Information requirement
<b>Cuff centre at right atrium level</b>	Yes	Information requirement
<b>Comfortable</b>	Yes	Information requirement
<b>5-min rest at start</b>	Yes	Information requirement
<b>Study Details Assessment</b>	<b>Checks</b>	15
	<b>Permitted Modifications</b>	0
	<b>Violations</b>	0

## PROCEDURE

### Screening and Recruitment Details

Screening and Recruitment				Assessment
Total Screened			<b>85</b>	Information requirement satisfactory
Total Excluded			<b>0</b>	Information requirement satisfactory
Device Failure		<b>0</b>		Information requirement satisfactory
Poor Quality Sounds		<b>0</b>		Information requirement satisfactory
Cuff Size Unavailable		<b>0</b>		Information requirement satisfactory
Observer Disagreement		<b>0</b>		Information requirement satisfactory
Bigeminy		<b>0</b>		Information requirement satisfactory
Trigeminy		<b>0</b>		Information requirement satisfactory
Isolated VPB		<b>0</b>		Information requirement satisfactory
Atrial Fibrillation		<b>0</b>		Information requirement satisfactory
Other Reasons*		<b>0</b>		Information requirement satisfactory
Total Recruited ( <i>85 Required</i> )			<b>85</b>	Value within requirements
*Explanation Summary				
<i>Not applicable</i>				Information requirement satisfactory
<b>On Treatment Ranges (On Rx / Total)</b>				
SBP	Low	≤ 100 mmHg	<b>0 / 11</b>	Supplementary Information
	Medium	131 – 139 mmHg	<b>10 / 30</b>	Supplementary Information
	High	≥ 140 mmHg	<b>23 / 44</b>	Supplementary Information
DBP	Low	< 80 mmHg	<b>0 / 8</b>	Supplementary Information
	Medium	80 – 100 mmHg	<b>12 / 32</b>	Supplementary Information
	High	> 100 mmHg	<b>21 / 45</b>	Supplementary Information
<b>Screening and Recruitment Details Assessment</b>	<b>Checks</b>			13
	<b>Permitted Modifications</b>			0
	<b>Violations</b>			0

**Subject Details**

	<i>Requirement</i>	<i>Value</i>	<i>Assessment</i>	
Sex				
Male:Female	26..59 (≥ 30%)	<b>34:51</b>	Value within requirements	Value within requirements
Age (years)				
Range (Low:High)	≥ 13	<b>24:93</b>	Value within requirements	Value within requirements
Mean (SD)		<b>72.3 (14.2)</b>	Desirable Information	Desirable Information
Adults:Children	85:0	<b>85:0</b>	Value within requirements	Value within requirements
Arm Circumference (cm)				
Range (Low:High)		<b>20:37</b>	Desirable Information	Desirable Information
Mean (SD)		<b>28.5 (5.4)</b>	Desirable Information	Desirable Information
Cuff for Test Device (cm)				
Paediatric (17 – 22)	15 to 55 (≥ 16.7%)	<b>24 (28.2%)</b>	Value within requirements	
Standard (22 – 32)	15 to 55 (≥ 16.7%)	<b>35 (41.2%)</b>	Value within requirements	
Large (22 – 42)	15 to 55 (≥ 16.7%)	<b>26 (30.6%)</b>	Value within requirements	
Total	85	<b>85</b>	Value within requirements	
Recruitment SBP (mmHg)				
Range (Low:High)		<b>87:183</b>	Supplementary Information	Supplementary Information
Mean (SD)		<b>137.4 (24.5)</b>	Supplementary Information	Supplementary Information
Recruitment DBP (mmHg)				
Range (Low:High)		<b>39:125</b>	Supplementary Information	Supplementary Information
Mean (SD)		<b>86.6 (16.9)</b>	Supplementary Information	Supplementary Information

<b>Subject Details Assessment</b>	<b>Checks</b>	10
	<b>Permitted Modifications</b>	0
	<b>Violations</b>	0

**Observer Measurements Range-Requirements**

	<i>Requirement</i>	<i>Value</i>	<i>Assessment</i>	
SBP (mmHg)				
≤ 100	13 to 204 (≥ 5%, < 80%)	<b>40 (15.7%)</b>	Value within requirements	
101 to 139	0 to 191 (< 75%)	<b>97 (38.0%)</b>	Value within requirements	
≥ 140	51 to 242 (≥ 20%, < 95%)	<b>118 (46.3%)</b>	Value within requirements	
≥ 160	13 to 242 (≥ 5%, < 95%)	<b>49 (19.2%)</b>	Value within requirements	
DBP (mmHg)				
≤ 60	13 to 204 (≥ 5%, < 80%)	<b>23 (9.0%)</b>	Value within requirements	
61 to 84	0 to 191 (< 75%)	<b>120 (47.1%)</b>	Value within requirements	
≥ 85	51 to 242 (≥ 20%, < 95%)	<b>112 (43.9%)</b>	Value within requirements	
≥ 100	13 to 242 (≥ 5%, < 95%)	<b>49 (22.0%)</b>	Value within requirements	
Total		<b>255 (100%)</b>		

DBP sounds used				
K4:K5	Total 85	<b>0:85</b>	Information req. satisfactory	Information req. satisfactory

<b>Observer Measurements Range Assessment</b>	<b>Checks</b>	10
	<b>Permitted Modifications</b>	0
	<b>Violations</b>	0

**STUDY RESULTS**

**Observer Differences**

			<b>Assessment</b>	
Observer 2 – Observer 1				
SBP (mmHg)	Range (Low:High)	<b>-4:4</b>	Supplementary Information	Supplementary Information
	Mean (SD)	<b>0.7 (1.8)</b>	Supplementary Information	Supplementary Information
DBP (mmHg)	Range (Low:High)	<b>-4:4</b>	Supplementary Information	Supplementary Information
	Mean (SD)	<b>0.4 (1.9)</b>	Supplementary Information	Supplementary Information
Repeated Measurements		<b>1</b>	Information requirement satisfactory	
<b>Observer difference &gt; 4 mmHg</b>				


<b>Observer Differences Assessment</b>	<b>Checks</b>	1
	<b>Permitted Modifications</b>	0
	<b>Violations</b>	0


**Validation Results**

<b>Criterion 1</b>	Pass Req.	Achieved		<b>Assessment</b>	
		SBP	DBP		
Measurement pairs		<b>255</b>		Value within requirements	
Mean mmHg	≤ 5	<b>+3.6</b>	<b>+3.0</b>	Value within passing criteria	Value within passing criteria
SD mmHg	≤ 8	<b>2.7</b>	<b>2.7</b>	Value within passing criteria	Value within passing criteria
Criterion 1 Result		<b>Pass</b>	<b>Pass</b>	Value within passing criteria	Value within passing criteria
<b>Criterion 2</b>					
Number of subjects		<b>85</b>		Value within requirements	
Mean mmHg		<b>+3.6</b>	<b>+3.0</b>	Value within passing criteria	Value within passing criteria
SD mmHg	≤ 5.89:6.25	<b>2.6</b>	<b>2.6</b>	Value within passing criteria	Value within passing criteria
Criterion 2 Result		<b>Pass</b>	<b>Pass</b>	Value within passing criteria	Value within passing criteria
<b>Result</b>		<b>Pass</b>		Value within passing criteria	

<b>Validation Results Assessment</b>	<b>Checks</b>	15
	<b>Permitted Modifications</b>	0
	<b>Violations</b>	0

**Plots**





See Page 6 for Plot Legend

		<b>Assessment</b>
SBP Plot Provided	<b>Yes</b>	Information requirement satisfactory
DBP Plot Provided	<b>Yes</b>	Information requirement satisfactory
<b>Plots Assessment</b>		<b>Checks</b> 2 <b>Permitted Modifications</b> 0 <b>Violations</b> 0
<b>AAMI/ANSI/ISO 81060-2:2013 Primary Study Assessment</b>		<b>Checks</b> 66 <b>Permitted Modifications</b> 0 <b>Violations</b> 0

## AAMI/ANSI/ISO 81060-2:2013 Additional Requirements for ABPM Device

### STUDY DETAILS

<b>Protocol</b>	The AAMI/ANSI/ISO 81060-2:2013 additional requirements for a sphygmomanometer intended for use in ambulatory monitoring <sup>1</sup>	
		<b>Assessment</b>
<b>Reference Determination</b>	Simultaneous opposite-arm	Information requirement satisfactory
<b>Adherence</b>	Followed Precisely	Information requirement satisfactory
<b>Adjustments</b>	None	Information requirement satisfactory
<b>Study Meas. Method</b>	Auscultatory	Information requirement satisfactory
<b>Study Measurement Site</b>	Upper Arm	Information requirement satisfactory
<b>Observers</b>		
<b>Supervisor + 2 Observers</b>	Yes	Information requirement satisfactory
<b>Observer Training</b>	By expert in BP measurement	Information requirement satisfactory
<b>Observer Familiarisation</b>	40 test measurements	Supplementary Information
<b>Observers Blinded</b>	From device and each other	Information requirement satisfactory
<b>Sample</b>		
<b>Population</b>	A general population	Information requirement satisfactory
<b>Circumstances</b>	Bicycle ergometer stress for ABPM additional requirements	Information requirement satisfactory
<b>HBP Subjects Selection</b>	Inpatients and outpatients	Supplementary Information
<b>NBP Subjects Selection</b>	Healthcare staff	Supplementary Information
<b>Study Details Assessment</b>	<b>Checks</b>	10
	<b>Permitted Modifications</b>	0
	<b>Violations</b>	0

### PROCEDURE

#### Screening and Recruitment Details

Screening and Recruitment				Assessment
Total Screened			<b>36</b>	Information requirement satisfactory
Total Excluded			<b>0</b>	Information requirement satisfactory
Device Failure		<b>0</b>		Information requirement satisfactory
Poor Quality Sounds		<b>0</b>		Information requirement satisfactory
Cuff Size Unavailable		<b>0</b>		Information requirement satisfactory
Observer Disagreement		<b>0</b>		Information requirement satisfactory
Bigeminy		<b>0</b>		Information requirement satisfactory
Trigeminy		<b>0</b>		Information requirement satisfactory
Isolated VPB		<b>0</b>		Information requirement satisfactory
Atrial Fibrillation		<b>0</b>		Information requirement satisfactory
Other Reasons*		<b>0</b>		Information requirement satisfactory
Total Recruited ( $\geq 85$ Required)			<b>36</b>	Value within requirements
*Explanation Summary				Information requirement satisfactory
<i>Not applicable</i>				
On Treatment Ranges (On Rx / Total)				
Rest SBP	Low	$\leq 100$ mmHg	<b>1 / 2</b>	Supplementary Information
	Medium	131 – 139 mmHg	<b>8 / 19</b>	Supplementary Information
	High	$\geq 140$ mmHg	<b>6 / 15</b>	Supplementary Information
Rest DBP	Low	$< 80$ mmHg	<b>1 / 2</b>	Supplementary Information
	Medium	80 – 100 mmHg	<b>11 / 24</b>	Supplementary Information
	High	$> 100$ mmHg	<b>3 / 10</b>	Supplementary Information
<b>Screening and Recruitment Details Assessment</b>	<b>Checks</b>		13	
	<b>Permitted Modifications</b>		0	
	<b>Violations</b>		0	

**Subject Details**

	<i>Requirement</i>	<i>Value</i>	<i>Assessment</i>	
Sex				
Male:Female	≥ 11 (30%)	<b>14:22</b>	Value within requirements	Value within requirements
Age (years)				
Range (Low:High)	≥ 13	<b>28:88</b>	Value within requirements	Value within requirements
Mean (SD)		<b>70.8 (14.6)</b>	Desirable Information	Desirable Information
Adults:Children	36:0	<b>36:0</b>	Value within requirements	Value within requirements
Arm Circumference (cm)				
Range (Low:High)		<b>20:36</b>	Desirable Information	Desirable Information
Mean (SD)		<b>28.2 (5.7)</b>	Desirable Information	Desirable Information
Cuff for Test Device (cm)				
Paediatric (17 – 22)	6 to 24 (≥ 16.7%)	<b>12 (33.3%)</b>	Value within requirements	
Standard (22 – 32)	6 to 24 (≥ 16.7%)	<b>13 (36.1%)</b>	Value within requirements	
Large (22 – 42)	6 to 24 (≥ 16.7%)	<b>11 (30.6%)</b>	Value within requirements	
Total	36	<b>36</b>	Value within requirements	
Resting SBP (mmHg)				
Range (Low:High)		<b>86:166</b>	Desirable Information	Desirable Information
Mean (SD)		<b>134.2 (19.8)</b>	Desirable Information	Desirable Information
> 160 mmHg	4 to 36 (≥ 10%)	<b>6</b>	Value within requirements	
Resting DBP (mmHg)				
Range (Low:High)		<b>56:106</b>	Desirable Information	Desirable Information
Mean (SD)		<b>80.1 (13.8)</b>	Desirable Information	Desirable Information
> 100 mmHg	4 to 36 (≥ 10%)	<b>6</b>	Value within requirements	
<b>Subject Details Assessment</b>			<b>Checks</b>	12
			<b>Permitted Modifications</b>	0
			<b>Violations</b>	0

**Stress Details**

	<i>Requirement</i>	<i>Value</i>	<i>Assessment</i>	
Resting Pulse Rate (bpm)				
Range (Low:High)		<b>62:88</b>	Information req. satisfactory	Information req. satisfactory
Mean (SD)		<b>73.1 (6.5)</b>	Information req. satisfactory	Information req. satisfactory
Bicycle Ergometer Percentage Load Settings				
Range (Low:High)		<b>11:18</b>	Information req. satisfactory	Information req. satisfactory
Mean (SD)		<b>14.2 (1.8)</b>	Information req. satisfactory	Information req. satisfactory
Stressed Pulse Rate (bpm)				
Range (Low:High)		<b>70:102</b>	Information req. satisfactory	Information req. satisfactory
Mean (SD)		<b>83.4 (7.9)</b>	Information req. satisfactory	Information req. satisfactory
Pulse Rate Percentage Increase				
Range (Low:High)	≥ 10:≤ 20	<b>10.9:17.6</b>	Value within requirements	Value within requirements
Mean (SD)		<b>14.0 (1.8)</b>	Information req. satisfactory	Information req. satisfactory
< 10%	0	<b>0</b>	Value within requirements	
10% to 20%	36	<b>36</b>	Value within requirements	
> 20%	0	<b>0</b>	Value within requirements	
Correlation of PR Percentage Increase with Bicycle Ergometer Percentage Load				
r		<b>0.74</b>	Supplementary Information	
<b>Stress Details Assessment</b>			<b>Checks</b>	19
			<b>Permitted Modifications</b>	0
			<b>Violations</b>	0

### Observer Measurements Range-Requirements

	<i>Requirement</i>	<b>Value</b>	<b>Assessment</b>	
<i>SBP (mmHg)</i>				
	≤ 100	<b>6 (5.6%)</b>	Value within requirements	
	101 to 139	<b>55 (50.9%)</b>	Value within requirements	
	≥ 140	<b>47 (43.5%)</b>	Value within requirements	
	≥ 160	<b>18 (16.7%)</b>	Value within requirements	
<i>DBP (mmHg)</i>				
	≤ 60	<b>4 (3.7%)</b>	Value within requirements	
	61 to 84	<b>68 (63.0%)</b>	Value within requirements	
	≥ 85	<b>36 (33.3%)</b>	Value within requirements	
	≥ 100	<b>19 (17.6%)</b>	Value within requirements	
DBP sounds used				
K4:K5	Total 36	<b>0:36</b>	Information req. satisfactory	Information req. satisfactory

<b>Observer Measurements Range Assessment</b>	<b>Checks</b>	10
	<b>Permitted Modifications</b>	0
	<b>Violations</b>	0

### STUDY RESULTS

#### Observer Differences

			<b>Assessment</b>	
Observer 2 – Observer 1				
<i>SBP (mmHg)</i>	Range ( <i>Low:High</i> )	<b>-4:4</b>	Optional data provided	Optional data provided
	Mean (SD)	<b>-0.3 (1.9)</b>	Optional data provided	Optional data provided
<i>DBP (mmHg)</i>	Range ( <i>Low:High</i> )	<b>-4:4</b>	Optional data provided	Optional data provided
	Mean (SD)	<b>0.0 (1.9)</b>	Optional data provided	Optional data provided
Repeated Measurements		<b>2</b>	Information requirement satisfactory	
		<b>Low sound quality</b>		

<b>Observer Differences Assessment</b>	<b>Checks</b>	1
	<b>Permitted Modifications</b>	0
	<b>Violations</b>	0



#### Validation Results

	Pass Req.	Achieved		<b>Assessment</b>	
		SBP	DBP		
<b>Criterion 1</b>					
Measurement pairs		<b>108</b>		Value within requirements	
Mean <i>mmHg</i>	≤ 5	<b>+4.0</b>	<b>+4.0</b>	Value within passing criteria	Value within passing criteria
SD <i>mmHg</i>	≤ 8	<b>3.3</b>	<b>3.4</b>	Value within passing criteria	Value within passing criteria
Criterion 1 Result		<b>Pass</b>	<b>Pass</b>	Value within passing criteria	Value within passing criteria
<b>Criterion 2</b>					
Number of subjects		<b>36</b>		Value within requirements	
Mean <i>mmHg</i>		<b>+4.0</b>	<b>+4.0</b>	Value within passing criteria	Value within passing criteria
SD <i>mmHg</i>	≤ 5.64:5.64	<b>3.1</b>	<b>3.3</b>	Value within passing criteria	Value within passing criteria
Criterion 2 Result		<b>Pass</b>	<b>Pass</b>	Value within passing criteria	Value within passing criteria
<b>Result</b>		<b>Pass</b>		Value within passing criteria	

<b>Validation Results Assessment</b>	<b>Checks</b>	15
	<b>Permitted Modifications</b>	0
	<b>Violations</b>	0



**Plots**



See Page 6 for Plot Legend

		<b>Assessment</b>
SBP Plot Provided	<b>Yes</b>	Information requirement satisfactory
DBP Plot Provided	<b>Yes</b>	Information requirement satisfactory
<b>Plots Assessment</b>		<b>Checks</b> 2 <b>Permitted Modifications</b> 0 <b>Violations</b> 0
<b>AAMI/ANSI/ISO 81060-2:2013 Additional Requirements for ABPM Device Assessment</b>		<b>Checks</b> 82 <b>Permitted Modifications</b> 0 <b>Violations</b> 0

## ESH-IP 2010 Study

### STUDY DETAILS

<b>Protocol</b>	The European Society of Hypertension International Protocol revision 2010 for the validation of blood pressure measuring devices in adults <sup>2</sup>	
		<b>Assessment</b>
<b>Adherence</b>	Followed Precisely	Information requirement satisfactory
<b>Adjustments</b>	None	Information requirement satisfactory
<b>Study Meas. Method</b>	Auscultatory	Information requirement satisfactory
<b>Study Measurement Site</b>	Upper Arm	Information requirement satisfactory
<b>Observers</b>		
<b>Supervisor + 2 Observers</b>	Yes	Information requirement satisfactory
<b>Observer Training</b>	By expert in BP measurement	Information requirement satisfactory
<b>Observer Familiarisation</b>	40 test measurements	Information requirement satisfactory
<b>Observers Blinding</b>	From device and each other	Information requirement satisfactory
<b>Sight and Hearing Checked</b>	Yes	Information requirement
<b>Sample</b>		
<b>Population</b>	A general population	Information requirement satisfactory
<b>Circumstances</b>	None	Information requirement satisfactory
<b>HBP Subjects Selection</b>	Inpatients and outpatients	Information requirement satisfactory
<b>NBP Subjects Selection</b>	Healthcare staff	Information requirement satisfactory
<b>Subject Preparation</b>		
<b>Back, elbow and arm supported</b>	Yes	Information requirement
<b>Legs uncrossed</b>	Yes	Information requirement
<b>Cuff centre at right atrium level</b>	Yes	Information requirement
<b>10-15-min rest at start</b>	Yes	Information requirement
<b>Study Details Assessment</b>	<b>Checks</b>	17
	<b>Permitted Modifications</b>	0
	<b>Violations</b>	0

**PROCEDURE**

**Table 1: Screening and Recruitment Details**

<b>Screening and Recruitment</b>				<b>Assessment</b>
Total Screened			<b>35</b>	Information requirement satisfactory
Total Excluded			<b>2</b>	Information requirement satisfactory
	Ranges Complete		<b>2</b>	Information requirement satisfactory
	Range Adjustment		<b>0</b>	Information requirement satisfactory
	Arrhythmias		<b>0</b>	Information requirement satisfactory
	Device Failure		<b>0</b>	Information requirement satisfactory
	Poor Quality Sounds		<b>0</b>	Information requirement satisfactory
	Cuff Size Unavailable		<b>0</b>	Information requirement satisfactory
	Observer Disagreement		<b>0</b>	Information requirement satisfactory
	Distribution		<b>0</b>	Information requirement satisfactory
	Other Reasons*		<b>0</b>	Information requirement satisfactory
Total Recruited (33 Required)			<b>33</b>	Value within requirements
*Explanation Summary				Information requirement satisfactory
Not applicable				

		<b>Recruitment Ranges</b>		<b>Assessment</b>
		<b>Req.</b>	<b>Values</b>	
<b>SBP</b>				
Total		33	<b>33</b>	Value within requirements
Low	< 130 mmHg	10..12	<b>10</b>	Value within requirements
	< 90 mmHg		<b>1</b>	Information requirement satisfactory
	90 – 129 mmHg		<b>9</b>	Information requirement satisfactory
Medium	130 – 160 mmHg	10..12	<b>12</b>	Value within requirements
High	> 160 mmHg	10..12	<b>11</b>	Value within requirements
	161 – 180 mmHg		<b>10</b>	Information requirement satisfactory
	> 180 mmHg		<b>1</b>	Information requirement satisfactory
<b>DBP</b>				
Total		33	<b>33</b>	Value within requirements
Low	< 80 mmHg	10..12	<b>10</b>	Value within requirements
	< 40 mmHg		<b>1</b>	Information requirement satisfactory
	40 – 79 mmHg		<b>9</b>	Information requirement satisfactory
Medium	80 – 100 mmHg	10..12	<b>12</b>	Value within requirements
High	> 100 mmHg	10..12	<b>11</b>	Value within requirements
	101 – 130 mmHg		<b>11</b>	Information requirement satisfactory
	> 130 mmHg		<b>0</b>	Information requirement satisfactory
Total Extremes		0..4	<b>3</b>	Value within requirements

<b>On Treatment Ranges</b>			
SBP	Low	< 130 mmHg	<b>1</b>
	Medium	130 – 160 mmHg	<b>5</b>
	High	> 160 mmHg	<b>8</b>
DBP	Low	< 80 mmHg	<b>1</b>
	Medium	80 – 100 mmHg	<b>5</b>
	High	> 100 mmHg	<b>8</b>

<b>Screening and Recruitment Details Assessment</b>	<b>Checks</b>	36
	<b>Permitted Modifications</b>	0
	<b>Violations</b>	0

**Table 2: Subject Details**

	<i>Requirement</i>	<i>Value</i>	<i>Assessment</i>	
Sex				
Male:Female	≥ 10: ≥ 10	<b>14:19</b>	Value within requirements	Value within requirements
Age (years)				
Range (Low:High)	≥ 25	<b>59:93</b>	Value within requirements	Information req. satisfactory
Mean (SD)		<b>76.8 (7.3)</b>	Information req. satisfactory	Information req. satisfactory
Arm Circumference (cm)				
Range (Low:High)		<b>20:37</b>	Information req. satisfactory	Information req. satisfactory
Mean (SD)		<b>28.8 (5.2)</b>	Information req. satisfactory	Information req. satisfactory
Cuff for Test Device (cm)				
Paediatric Plus (17 – 22)		<b>8</b>	Information requirement satisfactory	
Standard Plus (22 – 32)		<b>15</b>	Information requirement satisfactory	
Large Plus (22 – 42)		<b>10</b>	Information requirement satisfactory	
Total	33	<b>33</b>	Value within requirements	
Recruitment SBP (mmHg)				
Range (Low:High)		<b>89:183</b>	Information req. satisfactory	Information req. satisfactory
Mean (SD)		<b>140.8 (28.8)</b>	Information req. satisfactory	Information req. satisfactory
Recruitment DBP (mmHg)				
Range (Low:High)		<b>39:125</b>	Information req. satisfactory	Information req. satisfactory
Mean (SD)		<b>90.1 (18.8)</b>	Information req. satisfactory	Information req. satisfactory
<b>Subject Details Assessment</b>			<b>Checks</b>	22
			<b>Permitted Modifications</b>	0
			<b>Violations</b>	0

**Table 3: Observer Measurements in each Recruitment Range**

	<i>Requirement</i>	<i>Value</i>	<i>Assessment</i>	
SBP				
Overall Range mmHg (Low:High)	≤ 100: ≥ 170	<b>85:182</b>	Value within requirements	Value within requirements
Low (< 130 mmHg)	22..44	<b>33</b>	Value within requirements	
Medium (130 – 160 mmHg)	22..44	<b>38</b>	Value within requirements	
High (> 160 mmHg)	22..44	<b>28</b>	Value within requirements	
Maximum Difference	≤ 19	<b>10</b>	Value within requirements	
DBP				
Overall Range mmHg (Low:High)	≤ 50: ≥ 120	<b>35:121</b>	Value within requirements	Value within requirements
Low (< 80 mmHg)	22..44	<b>32</b>	Value within requirements	
Medium (80 – 100 mmHg)	22..44	<b>38</b>	Value within requirements	
High (> 100 mmHg)	22..44	<b>29</b>	Value within requirements	
Maximum Difference	≤ 19	<b>9</b>	Value within requirements	
<b>Observer Measurements Range Assessment</b>			<b>Checks</b>	12
			<b>Permitted Modifications</b>	0
			<b>Violations</b>	0

**STUDY RESULTS**



**Table 4: Observer Differences**

	<b>Requirement</b>	<b>Value</b>	<b>Assessment</b>	
SBP ( <i>mmHg</i> ) Observer 2 – Observer 1				
Range ( <i>Low:High</i> )	-4..+4: -4..+4	<b>-2:+4</b>	Value within requirements	Value within requirements
Mean (SD)		<b>+1.3 (1.4)</b>	Information req. satisfactory	Information req. satisfactory
DBP ( <i>mmHg</i> ) Observer 2 – Observer 1				
Range ( <i>Low:High</i> )	-4..+4: -4..+4	<b>-4:+4</b>	Value within requirements	Value within requirements
Mean (SD)		<b>+0.7 (1.7)</b>	Information req. satisfactory	Information req. satisfactory
Repeated Measurements		<b>0</b>	Information requirement satisfactory	
<b>Observer Differences Assessment</b>			<b>Checks</b>	9
			<b>Permitted Modifications</b>	0
			<b>Violations</b>	0

**Table 5: Validation Results**

<b>Part 1</b>	Pass Req.		Achieved		<b>Assessment</b>	
	Two of	All of	SBP	DBP		
$\leq 5$ <i>mmHg</i>	73	65	<b>94</b>	<b>94</b>	Value within upper criteria	Value within upper criteria
$\leq 10$ <i>mmHg</i>	87	81	<b>99</b>	<b>99</b>	Value within upper criteria	Value within upper criteria
$\leq 15$ <i>mmHg</i>	96	93	<b>99</b>	<b>99</b>	Value within upper criteria	Value within upper criteria
Grade 1			<b>Pass</b>	<b>Pass</b>	Value within upper criteria	Value within upper criteria
Mean <i>mmHg</i>			<b>3.4</b>	<b>2.9</b>	Information req. satisfactory	Information req. satisfactory
SD <i>mmHg</i>			<b>2.1</b>	<b>2.3</b>	Information req. satisfactory	Information req. satisfactory
<b>Part 2</b>		Pass Req.	Achieved			
			SBP	DBP		
$2/3 \leq 5$ <i>mmHg</i>		$\geq 24$	<b>32</b>	<b>33</b>	Value within passing criteria	Value within passing criteria
$0/3 \leq 5$ <i>mmHg</i>		$\leq 3$	<b>0</b>	<b>0</b>	Value within passing criteria	Value within passing criteria
Grade 2			<b>Pass</b>	<b>Pass</b>	Value within passing criteria	Value within passing criteria
Grade 3			<b>Pass</b>	<b>Pass</b>	Value within passing criteria	Value within passing criteria
<b>Part 3</b>			<b>Pass</b>		All values within upper passing criteria	
Result						
<b>Validation Results Assessment</b>			<b>Checks</b>	21		
			<b>Permitted Modifications</b>	0		
			<b>Violations</b>	0		

**Plots**



See Page 6 for Plot Legend

		<b>Assessment</b>
SBP Plot Provided	<b>Yes</b>	Information requirement satisfactory
DBP Plot Provided	<b>Yes</b>	Information requirement satisfactory
<b>Plots Assessment</b>		<b>Checks</b> 2 <b>Permitted Modifications</b> 0 <b>Violations</b> 0
<b>ESH Study Assessment</b>		<b>Checks</b> 119 <b>Permitted Modifications</b> 0 <b>Violations</b> 0

## Limitations

		<u>Assessment</u>
<b>Effect of Problems</b>	No problems	Information requirement satisfactory
<b>Justification of Adjustments</b>	No adjustments to protocol	Information requirement satisfactory
<b>Effect of K4 use</b>	K4 not used	Information requirement satisfactory
<b>Previous Validation Studies</b>	None	Information requirement satisfactory
<b>Comparisons</b>	Not applicable	Information requirement satisfactory
<b>Contrasts</b>	Not applicable	Information requirement satisfactory
<b>Cautions for correct use</b>	None	Information requirement satisfactory

<b>Limitations Assessment</b>	<b>Checks</b>	7
	<b>Permitted Modifications</b>	0
	<b>Violations</b>	0

## Recommendations

### Overall Summary

<i>Number of checks</i>	295
<i>Number of permitted modifications</i>	0
<i>Number of violations</i>	0

### Assessment Summary

The validations have been checked and are verified as having been conducted in accordance with the respective protocol requirements. Therefore, any hypothesis that the reliability of the results may be compromised due to protocol adjustment or violation must be rejected and each of the results must be considered to be valid. According to each protocol, these results require that the null hypothesis, that the tested technology in the device is inaccurate in measuring blood pressure, is rejected. Therefore, it must be concluded, that the auscultatory blood pressure measurement technology used in the Novacor Diasys 3 Plus monitor is accurate for blood pressure measurement in adults, in both static and ambulatory settings.

### Certification Decision

The Novacor Diasys 3 Plus (DIP-0001-00), with the appropriate Paediatric Plus 18 cm to 24 cm (ACC-0212-00), Standard Plus 24 cm to 32 cm (ACC-0210-00) or Large Plus 32 cm to 40 cm (ACC-0211-00) cuff, is certified by Medaval Ltd., for auscultatory blood pressure measurement, including ABPM, in adults, as the technology fulfilled the conditions required for a pass in 1) a primary validation study carried out in accordance with the requirements of the AAMI/ANSI/ISO 81060-2:2013 standard, 2) an additional validation study during stress carried out in accordance with the requirements of the AAMI/ANSI/ISO 81060-2:2013 standard for ambulatory monitors and 3) a validation study carried out in accordance with the requirements of the International Protocol of the European Society of Hypertension 2010 Revision.

Date of Approval: 04 March 2019.

## References

1. Association for the Advancement of Medical Instrumentation, American National Standards Institute, International Organization for Standardization. AAMI/ANSI/ISO 81060-2:2013, Non-invasive Sphygmomanometers - Part 2: Clinical Investigation of Automated Measurement Type. Geneva, Switzerland: ISO; 2013.
2. O'Brien E, Atkins N, Stergiou G, Karpettas N, Parati G, Asmar R, Imai Y, Wang J, Mengden T, Shennan A; Working Group on Blood Pressure Monitoring of the European Society of Hypertension. European Society of Hypertension International Protocol revision 2010 for the validation of blood pressure measuring devices in adults. *Blood Press Monit.* 2010;15:23-38. doi: 10.1097/MBP.0b013e3283360e98. PMID: 20110786. Erratum in *Blood Press Monit.* 2010;15(3):171-2.
3. European Commission – Health Technology and Cosmetics. MEDDEV 2.7/1 rev.4: Guidelines on Medical Devices – Clinical Evaluation: A Guide for Manufacturers and Notified Bodies under Directives 93/42/EEC and 90/385/EEC. Brussels, Belgium: European Commission; June 2016 (65 p). Available from: <http://ec.europa.eu/DocsRoom/documents/17522/attachments/1/translations/>
4. The European Parliament and the Council of the European Union. Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices, amending Directive 2001/83/EC, Regulation (EC) No 178/2002 and Regulation (EC) No 1223/2009 and repealing Council Directives 90/385/EEC and 93/42/EEC (Text with EEA relevance). Official Journal of the European Union. 2017 May 05;60(L 117):1-175. Available from: <https://eur-lex.europa.eu/legal-content/EN/TXT/?uri=CELEX%3A32017R0745>.