

ENGLISH

Dear customer,
Thank you for choosing one of our products. Our name stands for high-quality, thoroughly tested products for applications in the areas of heat, weight, blood pressure, body temperature, pulse, gentle therapy, massage, beauty, baby and air. Please read these instructions for use carefully and keep them for later use, be sure to make them accessible to other users and observe the information they contain.

With kind regards,
Your Beurer team

1. Included in delivery

1x PO 30 pulse oximeter, 2x 1.5 V LR03 AAA batteries, 1x Lanyard, 1x Belt bag, 1x These instructions for use

2. Intended use

Only use the Beurer PO 30 pulse oximeter on humans to measure the arterial oxygen saturation (SpO₂) of haemoglobin and the heart rate (pulse rate). The pulse oximeter is suitable for private use (at home) as well as for use in the medical sector (hospitals, medical establishments).

3. Getting to know your device

The Beurer PO 30 pulse oximeter provides a non-invasive measurement of the arterial oxygen saturation (SpO₂) and the heart rate (pulse rate). Oxygen saturation indicates the percentage of haemoglobin in arterial blood that is loaded with oxygen. Therefore it is an important parameter for as-sessing the respiratory function. To take a measurement, the pulse oximeter uses two rays of light with differing wavelengths, which strike the finger inserted inside the housing. A low oxygen saturation value generally indicates underlying illnesses (respiratory diseases, asthma, heart failure etc.). People with a low oxygen saturation value are more likely to experience the following symptoms: shortness of breath, increased heart rate, weak-ness, nervousness and outbreaks of sweating. If oxygen saturation is known to be chronically diminished, it requires monitoring using the pulse ox-i-meter under medical supervision. If you have acutely diminished oxygen saturation, with or without the accompanying symptoms, you must consult a doctor immediately as it could lead to a life-threatening situation. The pulse oximeter is particularly suitable for patients at risk such as people with heart disease or asthma, but also for athletes and healthy people who exercise at high altitude (e.g. mountaineers, skiers or amateur pilots).

4. Signs and symbols

The following symbols are used in these instructions for use, on the packaging and on the type plate for the device:

	WARNING Warning instruction indicating a risk of injury or dam-age to health		Permissible storage temperature and humidity
	IMPORTANT Safety note regarding potential for damage to the device/accessories		Permissible operating temperature and humidity
	Note Note on important information		Application part, type BF
	Observe the instructions for use		Serial number
	Arterial oxygen saturation of haemoglobin (in percent)		The CE labelling certifies that the product complies with the essential requirements of Directive 93/42/EEC on medical products.
	Pulse rate (beats per minute)		Manufacturer
	Disposal in accordance with EC Directive WEEE (Waste Electrical and Electronic Equipment)		Alarm suppression
	Do not dispose of batteries containing hazardous sub-stances with household waste.		Device protected against foreign objects ≥ 12.5 mm and against falling drops of water

5. Warnings and safety notes

Read these instructions for use carefully. Non-observance of the following information may result in personal injury or material damage. Store these instruc-tions for use and make them accessible to other users. Make sure you include these instructions for use when handing over the device to third parties.

⚠ WARNING

- Check to ensure that the package contains all the parts that should be included in the delivery.
- Check the pulse oximeter regularly before use to ensure that there is no visible damage to the device and the batteries are still sufficiently charged. In case of doubt, do not use the device and contact Beurer customer services or an authorised retailer.
- Do not use any additional parts that are not recommended by the manufacturer or offered as equipment.
- Under no circumstances should you open or repair the device yourself, as faultless functionality could no longer be guaranteed thereafter. Failure to comply will result in voiding of the warranty. For repairs, please contact Beurer customer services or an authorised retailer.

Do NOT use the pulse oximeter

- if you are allergic to rubber products.
- if the device or the finger you are using is damp.
- on small children or babies.
- during an MRI or CT scan.
- whilst taking a blood pressure measurement on the same arm using a cuff.
- on fingers that have nail varnish on, are dirty or have a plaster or other dressing on them.
- on large fingers that do not fit into the device easily (fingertip: width approx. > 20 mm, thickness >15 mm).
- on fingers with anatomical changes, oedemas, scars or burns.
- on fingers that are too small, as with small children for example (width approx. < 10 mm, thickness < 5 mm).
- on patients who are not steady at the site of application (e.g. trembling).
- near flammable or explosive gas mixtures.

- Using the device for long periods may cause pain for people with circulatory disorders. Therefore do not use the pulse oximeter for longer than approx. 2 hours on one finger.
- The pulse oximeter displays a current measurement but cannot be used for continuous monitoring.
- The pulse oximeter does not have an alarm function and is therefore not suitable for evaluating medical results.
- Do not self-diagnose or self-medicate on the basis of the measurements without consulting your doctor. In particular, do not start taking any new medication or change the type and/or dosage of any existing medication without prior approval.
- Do not look directly inside the housing during the measurement. The red light and the invisible infra-red light in the pulse oximeter are harmful to your eyes.
- This device is not intended for use by people (including children) with restricted physical, sensory or mental skills or a lack of experience and/ or a lack of knowledge, unless they are supervised by a person who has responsibility for their safety or they receive instructions from this person on how to use the device. Children should be supervised around the device to ensure they do not play with it.
- Neither of the displays for the pulse wave and pulse bar allows the strength of the pulse or circulation to be evaluated at the measurement site. Rather, they are exclusively used to display the current visual signal variation at the measurement site and do not enable reliable diagnostics for the pulse.


Non-observance of the following instructions can lead to incorrect or failed measurements.

- There must not be any nail varnish, artificial nails or other cosmetics on the finger to be measured.
- Ensure that the finger nail on the finger to be measured is short enough that the fingertip covers the sensor element in the housing.
- Keep your hand, finger and body steady during the measurement.
- For people with cardiac arrhythmia, the measurement values of SpO₂ and the heart rate may be incorrect or the measurement may not be possible at all.
- In cases of carbon monoxide poisoning, the pulse oximeter displays a measurement value that is too high.
- To avoid falsifying the measuring result, there should not be any strong light sources (e.g. fluorescent lamps or direct sunlight) in the immedi-ate vicinity of the pulse oximeter.
- People with low blood pressure, who suffer from jaundice or take medication for vascular contraction, may experience incorrect or falsified measurements.
- Incorrect measurements are likely for patients who have been administered medical dye in the past or for those who have abnormal haemo-globin levels. This applies in particular for cases of carbon monoxide poisoning and methaemoglobin poisoning, which can occur for example from the administration of local anaesthetics or from an existing methaemoglobin reductase deficiency.
- Protect the pulse oximeter from dust, shocks, moisture, extreme temperatures and explosive materials.

⚠ Notes on handling batteries

- If your skin or eyes come into contact with battery fluid, flush out the affected areas with water and seek medical assistance.
- ⚠ Choking hazard! Small children may swallow and choke on batteries. Store the batteries out of the reach of small children.
- Observe the plus (+) and minus (-) polarity signs.
- If a battery has leaked, put on protective gloves and clean the battery compartment with a dry cloth.
- Protect the batteries from excessive heat.
- ⚠ Risk of explosion! Never throw batteries into a fire.
- Do not charge or short-circuit batteries.
- If the device is not to be used for a long period, take the batteries out of the battery compartment.
- Use identical or equivalent battery types only.
- Always replace all batteries at the same time.
- Do not use rechargeable batteries.
- Do not disassemble, split or crush the batteries.


6. Unit description



7. Initial use

7.1 Inserting the batteries		
1. Slide the battery com-partment lid open.	2. Place the two batteries supplied in the pulse ox-i-meter as shown. Ensure that the correct battery polarity is observed.	3. Close the battery com-partment lid again.

Display description



7.2 Attaching the lanyard	
To transport the pulse oximeter more easily you can attach a lanyard to the device.	
1. Insert the narrow end of the lanyard through the holder as shown.	2. Draw the other end of the lanyard through the loop at the narrow end and tighten.

8. Operation

1. Insert one finger into the finger opening of the pulse oximeter as shown and hold it steady.	2. Press the function button. The pulse oxime-ter begins its measurement. Do not move during the measurement.	3. Your measurement values will appear on the screen after a few seconds.

ⓘ Note

When you remove your finger from the pulse oximeter, the device will automatically switch off after approx. five seconds.

Function button

The function button on the pulse oximeter has two functions in total:

- **Switch-on function:** When the pulse oximeter is switched off you can hold down the function button briefly to switch it on.
- **Brightness function:** To select your desired display brightness, hold down the function button for slightly longer during operation.

ⓘ Note:

The display orients automatically (vertical format, horizontal format). This ensures that the values are easy to read on the display at all times, regardless of how you hold the pulse oximeter.

9. Evaluating measurement results

⚠ WARNING		Decline in oxygen saturation depending on altitude		
The following table for evaluating your measurements does NOT apply to people with certain pre-existing conditions (e.g. asthma, heart failure, respiratory diseases) or whilst staying at altitudes above 1500 metres. If you have a pre-existing condition, always consult your doctor to evaluate your measure-ments.		ⓘ Note The following table informs you of the effects of various altitudes on oxygen saturation value and its impact on the human body. The following table does NOT apply to people with certain pre-existing conditions (e.g. asthma, heart failure, respiratory diseases etc.). People with pre-existing conditions can show signs of illness (e.g. hypoxia) at lower altitudes.		
SpO ₂ measurement (oxygen saturation) in %	Classification/measures to be taken	Altitude	Expected SpO ₂ value (oxygen satu-ration) in %	Impact on human body
99-94	Normal range	1500-2500 m	> 90	No altitude sickness (normally)
93-90	Decreased range: visit to the doctor recommended	2500-3500 m	~90	Altitude sickness, acclima-tisation recommended
< 90	Critical range Seek medical attention urgently	3500-5800 m	<90	Very frequent altitude sick-ness, acclimatisation abso-lutely essential
		5800-7500 m	<80	Severe hypoxia, only limited length of stay possible
		7500-8850 m	<70	Immediate, acute danger to life

Source: Hackett PH, Roach RC: High-Altitude Medicine. In: Auerbach PS (ed): Wilder-ness Medicine, 3rd edition; Mosby, St.Louis, MO 1995; 1-37.

10. Maintenance/cleaning

⚠ IMPORTANT:

Do not use high-pressure sterilisation on the pulse oximeter!
Under no circumstances should you hold the pulse oximeter under water, as this can cause liquid to enter and damage the pulse oximeter.

- Clean the housing and the interior rubber surface with a soft cloth dampened with medical alcohol after each use.
- If a low battery status appears on the display of the pulse oximeter, change the batteries.
- If you are not going to use the pulse oximeter for more than one month, remove both batteries from the device to avoid possible leaking.

11. Storage

⚠ IMPORTANT:

Store the pulse oximeter in a dry place (relative humidity ≤95 %). If the humidity is too high it may shorten the service life of the pulse oximeter or damage it. Store the pulse oximeter in a place where the ambient temperature is between -40°C and 60°C.

12. Disposal

For environmental reasons, do not dispose of the device in the household waste at the end of its useful life. Dispose of the unit at a suit-able local collection or recycling point. Dispose of the device in accordance with EC Directive – WEEE (Waste Electrical and Electronic Equipment). If you have any questions, please contact the local authorities responsible for waste disposal. The empty, completely flat batteries must be disposed of through specially designated collection boxes, recycling points or electronics retailers. You are legally required to dispose of the batteries.

The codes below are printed on batteries containing harmful substances:

Pb = Battery contains lead,
Cd = Battery contains cadmium,
Hg = Battery contains mercury.

13. What if there are problems?

Problem	Possible cause	Solution
The pulse oximeter is not display-ing measurement values.	The batteries in the pulse oximeter are empty.	Replace the batteries.
	Batteries not inserted correctly.	Reinsert the batteries. If after reinserting the batter-ies correctly there are still no measurement values displayed, contact customer services.
The pulse oximeter is displaying measurement interruptions or high measurement value jumps.	Insufficient circulation in the measurement finger.	Observe the warnings and safety notes in chapter 5.
	Measurement finger is too large or too small.	Fingertip must have the following measurements: Width between 10 and 22 mm Thickness between 5 and 15 mm
	Finger, hand or body is moving.	Keep your finger, hand and body still during the measurement.
	Cardiac arrhythmia	Seek medical attention.

14. Technical Data

Model no.	PO 30
Measurement method	Non-invasive measurement of arterial oxygen saturation of haemoglobin and pulse rate in finger
Measurement range	SpO ₂ 0 – 100%, Pulse 30 – 250 beats/minute
Accuracy	SpO ₂ 70 –100%, ±2%, Pulse 30 – 250 bpm, ±2 beats/minute
Dimensions	L 61 mm x W 36 mm x H 32 mm
Weight	Approx. 58 g (including batteries)
Sensor to measure SpO ₂	Red light (wave length 660 nm); infra-red (wave length 880 nm); silicon receiver diode
Permissible operating conditions	+10 °C to +40 °C, <75 % relative humidity, 700–1060 hPa ambient pressure
Permissible storage conditions	-40 °C to +60 °C, ≤95 % relative humidity, 500–1060 hPa ambient pressure
Power supply	2x 1.5V — AAA batteries
Battery life	2 AAA batteries last for approx. 2 years of operation at 3 measurements per day (each of 60 seconds).
Classification	IP22, application part type BF

The serial number is located on the device or in the battery compartment.

Technical information is subject to change without notification to allow for updates.

- This device conforms with the European standards EN60601-1 and EN60601-1-2 (in accordance with CISPR 11, IEC 61000-4-2, IEC 61000-4-3, IEC 61000-4-8) and is subject to particular precautions with regard to electromagnetic compatibility. Please note that portable and mobile HF communication systems may interfere with this device. For more details, please contact our Customer Services at the address indicated.
- This device complies with the EU Directive 93/42/EEC concerning medical devices, the Medizinproduktegesetz (German Medical Devices Act) and the DIN EN ISO 80601-2-61 standard (Medical electrical equipment – Particular requirements for the basic safety and essential perfor-mance of pulse oximeter equipment for medical use)

⚠ Notes on electromagnetic compatibility

- The device is suitable for use in all environments listed in these instructions for use, including domestic environments.
- The use of the device may be limited in the presence of electromagnetic disturbances. This could result in issues such as error messages or the failure of the display/device.
- Avoid using this device directly next to other devices or stacked on top of other devices, as this could lead to faulty operation. If, however, it is necessary to use the device in the manner stated, this device as well as the other devices must be monitored to ensure they are working properly.
- The use of accessories other than those specified or provided by the manufacturer of this device can lead to an increase in electromagnetic emissions or a decrease in the device's electromagnetic immunity; this can result in faulty operation.
- Keep portable RF communication devices (including peripheral equipment, such as antenna cables or external antennas) at least 30 cm away from all device parts, including all cables included in delivery. Failure to comply with the above can impair the performance of the device.
- Failure to comply with the above can impair the performance of the device.

15. Warranty/service

In case of a claim under the warranty please contact your local dealer or the local representation which is mentioned in the list "service internati-onal".

In case of returning the unit please add a copy of your receipt and a short report of the defect.

The following warranty terms shall apply:

1. The warranty period for BEURER products is either 5 years or- if longer- the country specific warranty period from date of Purchase. In case of a warranty claim, the date of purchase has to be proven by means of the sales receipt or invoice.
2. Repairs (complete unit or parts of the unit) do not extend the warranty period.
3. The warranty shall not be valid for damages because of
 - a. improper treatment, e.g. nonobservance of the user instructions.
 - b. repairs or tampering by the customer or unauthorised third parties.
 - c. transport from the manufacturer to the consumer or during transport to the service centre.
 - d. The warranty shall not be valid for accessories which are subject to normal wear and tear (cuff, batteries etc.).
4. Liability for direct or indirect consequential losses caused by the unit are excluded even if the damage to the unit is accepted as a warran-ty claim.

Subject to errors and changes